



4G LTE

belle®

go with confidence!

mobile
medical
alert
system



HOW TO CHARGE BELLE

IMPORTANT: Fully charge **Belle** before first use.



Plug in the charging cradle and place **Belle** into the cradle as shown.

Typical charging time is 2-3 hours. The red battery light flashes every 5 seconds while charging and is solid when **Belle** is fully charged.

It is normal for the light around the call button on **Belle** to be off when the device is not in the charging cradle.

You can check the battery level at any time by pressing the battery level indicator button on the side of the device. The device will either say "Battery OK" or "Battery low, please charge."

NOTE: **Belle's** rechargeable battery lasts up to 30 days per charge. Only charge as needed. Wear your device as often as possible to be protected.

belle®



Red battery light at top of call button

While charging, the red battery light will flash every 5 seconds until the device is fully charged.

Once fully charged, the red battery light will be solid.

Blue call light around call button

During an SOS call, the blue call light will be solid.

The blue call light will flash every 5 seconds while charging and for a short time after a call ends.

The blue call light also flashes every 5 seconds for 3 minutes during the daily check-in to the network.

HOW TO CALL FOR HELP

1. Press and hold the call button and count 1, 2, 3, or until the light around the call button on **Belle**[®] turns blue, then release the button.
2. After a short delay, you will hear a voice message and then tones or ringing.
3. The specialist will answer the call.
4. The specialist can send emergency personnel or loved ones to help you.

TESTING BELLE WEEKLY

*We recommend that you test **Belle** weekly.*

1. Press and hold the call button and count 1, 2, 3, or until the light around the call button on **Belle**[®] turns blue, then release the button.
2. After a short delay, you will hear a voice message and then tones or ringing.
3. The specialist will answer the call.
4. Tell the specialist you are testing.

If your **Belle** does not test properly, please contact your dealer immediately.

REPLACE LANYARD WITH BELT CLIP

1. Remove the lanyard by pressing the small tab upward to release the clip.
2. Pull the lanyard and clip away from the **Belle**® device.
3. Press the clip on the back of the belt clip into the slot on the back of **Belle** until it clicks.

Note: You can charge **Belle** in its cradle with the belt clip attached.



CLEANING BELLE WEEKLY

*We recommend that you clean **Belle** weekly to ensure proper charging.*

Take a soft cloth, such as an eyeglass cleaning cloth, and gently rub the gold contacts on the device and charging cradle to remove dirt, oil and debris.



You can also spray household cleaner onto the cleaning cloth (though do not spray directly onto the device or cradle) to remove debris.

BY USING THIS DEVICE, YOU ACKNOWLEDGE

Coverage:

This product requires that there be adequate cellular coverage to work properly. It is important to test the device to know if it works in your area. Remember that your environmental and topographical conditions may also affect your coverage. If you experience coverage issues, please contact your dealer immediately.

Charging:

Your device's rechargeable battery may last up to 30 days per charge. Wear your device at all times and only charge when necessary. Failure to follow charging procedures will result in the device not being able to function properly.

Water-Resistant:

Belle[®] is IP67 water-resistant. **Belle** should not be submerged. **Belle** should be towel-dried after exposure to water.

Pacemakers:

Individuals with pacemakers should consult their physician and review their pacemaker materials regarding interaction with cell phones, and take the same precautions the materials recommend for this device.

AND ACCEPT THE FOLLOWING INFORMATION.

Location-Based Services:

Some of the Equipment uses technology to permit third-parties, including the Monitoring Center, to determine where you are physically located at any given time (the "Location Based Services"). Location Based Services may work even if you are not in communication with the Monitoring Center. The accuracy of the Location Based Services is limited, and the Company, the Operators, the Monitoring Center, the Responders, or others may not be able to identify your location or the location of the Equipment precisely or at all. You authorize the Company to collect location-based information. We will only share your location-based information with the Operators, the Monitoring Center, the Responders, any person or entity that acquires the Company and/or the Company's interest in and to the Monitoring Services, or any other person or entity you specifically designate and only for the purpose of providing and improving the Monitoring Services.

ANY AND ALL LOCATION-BASED INFORMATION IS MADE AVAILABLE FOR INFORMATIONAL AND PLANNING PURPOSES ONLY AND IS NOT INTENDED TO BE RELIED UPON IN SITUATIONS WHERE PRECISE LOCATION INFORMATION IS NEEDED OR WHERE ERRONEOUS, INACCURATE, TIME-DELAYED OR INCOMPLETE LOCATION OR MAP DATA MAY LEAD TO DEATH, PERSONAL INJURY, OR PROPERTY OR ENVIRONMENTAL DAMAGE. YOU AGREE THAT LOCATION-BASED INFORMATION MAY VARY FROM ACTUAL LOCATION(S), ROAD, OR TERRAIN CONDITIONS DUE TO FACTORS THAT CAN AFFECT THE ACCURACY OF THE MAP DATA, SUCH AS, BUT NOT LIMITED TO, WEATHER, ROAD, AND TRAFFIC CONDITIONS, GEOPOLITICAL EVENTS, AND CONDITION OR STATUS OF YOUR CELLULAR PHONE OR CELLULAR SERVICES. WE DO NOT GUARANTEE ACCURACY OR COMPLETENESS OF ANY LOCATION-BASED INFORMATION.

belle®